

Technical Laboratory Evaluation

Lab Reporting Form



Comprehensive product information can also be downloaded from www.hrsimon.com

Customer Name:	_____
Address:	_____
Phone:	_____
Facility Name:	_____

Date:	10/9/01 9:15
Samples:	From: _____
Customer Comments:	Q/C test-95 F @ 25"

Film:	Fuji	AD-M	AD-M
Chemical:		CKE	KODAK
pH:		10.4	10.38
Specific Gravity:		1.081	1.083

Speed Index:	1.45	1.45
Contrast Index:	2.21	2.26
Base + Fog:	0.18	0.18
Dmax:	3.94	3.9
Average Gradient:	3.25	3.17

Data used to determine Average gradient

St closest to(2+B+F):	2.26	2.22
On Step:	14	14
St closest to(.25+B+F)	0.31	0.32
On Step:	10	10
Step Difference:	4	4

1	0.18	0.18
2	0.18	0.18
3	0.18	0.18
4	0.18	0.18
5	0.18	0.18
6	0.19	0.18
7	0.2	0.19
8	0.21	0.21
9	0.24	0.24
10	0.31	0.32
11	0.53	0.51
12	0.86	0.87
13	1.45	1.45
14	2.26	2.22
15	2.74	2.77
16	3.21	3.19
17	3.48	3.48
18	3.64	3.69
19	3.8	3.83
20	3.89	3.88
21	3.94	3.9

cc: Mr. Robert Simon
Mr. Jason Simon

H & D Curves for Comparison

