

Technical Laboratory Evaluation

Lab Reporting Form

Comprehensive product information can also be downloaded from www.hrsimon.com



Customer Name: _____
 Address: _____
 Phone: _____
 Facility Name: _____

Date: 10/17/01 8:47
 Samples: _____ From: _____
 Customer Comments: Q/C test-95 F @ 25"

Film:	Konica	CMH	CMH
Chemical:		AGFA- HD	UHC
pH:		10.58	10.55
Specific Gravity:		1.0895	1.0815

Speed Index:		1.38	1.42
Contrast Index:		2.19	2.57
Base + Fog:		0.21	0.2
Dmax:		4.23	4.23
Average Gradient:		3.44	3.91

Data used to determine Average gradient

St closest to(2+B+F):		2.1	2.24
On Step:		13	13
St closest to(.25+B+F)		0.55	0.48
On Step:		10	10
Step Difference:		3	3

1	0.21	0.2
2	0.22	0.2
3	0.22	0.21
4	0.22	0.21
5	0.23	0.21
6	0.24	0.21
7	0.26	0.23
8	0.31	0.26
9	0.38	0.33
10	0.55	0.48
11	0.93	0.85
12	1.38	1.42
13	2.1	2.24
14	2.74	3.05
15	3.36	3.62
16	3.79	3.98
17	4.05	4.1
18	4.14	4.18
19	4.21	4.21
20	4.23	4.23
21	4.23	4.23

cc: Mr. Robert Simon
 Mr. Jason Simon

H & D Curves for Comparison

