

# Technical Laboratory Evaluation

## Lab Reporting Form

Comprehensive product information can also be downloaded from [www.hrsimon.com](http://www.hrsimon.com)



Customer Name: \_\_\_\_\_  
 Address: \_\_\_\_\_  
 Phone: \_\_\_\_\_  
 Facility Name: \_\_\_\_\_

Date: 10/17/01 8:47  
 Samples: \_\_\_\_\_ From: \_\_\_\_\_  
 Customer Comments: Q/C test-95 F @ 25"

Film:	Agfa	HDR-C	HDR-C
Chemical:		AGFA- HD	UHC
pH:		10.57	10.55
Specific Gravity:		1.1	1.0815

Speed Index:	1.36	1.15
Contrast Index:	1.86	1.89
Base + Fog:	0.32	0.26
Dmax:	4.42	4.41
Average Gradient:	3.10	3.15

Data used to determine Average gradient

St closest to(2+B+F):	2.47	2.38
On Step:	13	13
St closest to(.25+B+F)	0.61	0.49
On Step:	9	9
Step Difference:	4	4

1	0.32	0.26
2	0.32	0.27
3	0.32	0.27
4	0.33	0.28
5	0.34	0.28
6	0.36	0.3
7	0.4	0.33
8	0.47	0.38
9	0.61	0.49
10	0.87	0.71
11	1.36	1.15
12	1.85	1.68
13	2.47	2.38
14	2.99	3.01
15	3.37	3.43
16	3.7	3.76
17	3.97	3.98
18	4.19	4.18
19	4.32	4.34
20	4.35	4.41
21	4.42	4.41

cc: Mr. Robert Simon  
 Mr. Jason Simon

H & D Curves for Comparison

